

**Safety Communication**

**رسالة سلامة**

**Potential for invalid or false positive results.**

<b>Device/ Product Description:</b>	Analyzers, Laboratory, Immunoassay	
<b>Brand:</b>	Cobas z 480 Analyzer	
<b>Affected product:</b>	GMMI	05200881001
	Device Identifier	04015630929016
<b>Manufacturer:</b>	Roche Diagnostics GmbH.	
<b>Problem:</b>	Potential for false positive/invalid results due to dirty lenses. Factors that may contribute to dirty lenses are: improper sealing of the AD-plate, age and usage of the instrument, and laboratory environmental conditions.	
<b>Recommendation /Actions:</b>	<ul style="list-style-type: none"> <li>• Make sure that this document, with accompanying FSNs, are reached to the end-users.</li> <li>• The end-user should follow laboratory standards operation procedure to investigate the potential for false positive results for assays where a change in a result reporting could impact patient management.</li> <li>• Contact the Authorized Representative for required corrective action.</li> </ul> <p>For more information, Please click <a href="#">here</a>.</p> <p>If you think you had a problem with your device or a device your patient uses, please do not hesitate to report the problem to SFDA through:  <a href="#">NCMDR</a>  <a href="#">Vigilance system</a>  19999 unified call center</p>	

<b>Devices/Products photo:</b>		
<b>Authorized Representative Details</b>	AR name:	Roche Diagnostics Saudi Arabia Limited
	Assigned Contact Person:	Turki Abdulaziz Bin Juraid
	Mobile/Phone:	0555443544
	Email:	turki.juraid@roche.com

