



Medical Device Sector
Surveillance & Biometrics Executive Department

قطاع الأجهزة والمنتجات الطبية الإدارة التنفيذية للرقابة والقياسات الحيوية

Safety Communication

رسالة سلامة

May Be Susceptible to Network Session Vulnerability

Device/ Product Description:	Alaris 8015 PC Units and Alaris Systems Manager Software	
Manufacturer:	Becton, Dickinson and Company	
Problem Summary:	Network session vulnerability within the authentication process between the Specific PC units and the Specific software. If exploited, this vulnerability could allow an unauthorized user to establish a direct networking session between the two products. For more information, please check Here. If you think you had a problem with your device or a device your patient uses, please report the problem to SFDA through: NCMDR Vigilance system 19999 unified call center	
	AR name:	Becton Dickinson B.V.
	Assigned Contact Person: Mobile/Phone:	Saleh Al-Sohibanei 0544760333
	Email:	saleh al sohibanei@europe.bd.com

SG-2012-287-H 15/12/2020