

Safety Communication

رسالة سلامة

Sterility of the devices cannot be guaranteed

Device/ Product Description:	Infusion Sets for Alaris Pumps (GP, VP, CC and SE) & Non-Dedicated Infusion sets and accessories
Affected Lot Number:	Check the affected lot number in the original FIELD SAFETY NOTICE
Manufacturer:	BD
Problem:	The manufacturer is unable to guarantee the sterility of the devices. The use of non-sterile devices in the clinical setting could lead to an increased risk of infection which may cause serious harm or life-threatening conditions.
Recommendation /Actions:	<ul style="list-style-type: none">• Check stock for affected products and Cease use of all impacted lot numbers in your possession• Make sure that this document is reached to the end-users.• Follow instructions provided by the manufacturer in the FSN.• Contact the authorized representative for required support. <p>For more information, Please click here.</p> <p>If you think you had a problem with your device or a device your patient uses, please report the problem to SFDA through: NCMDR Vigilance system (19999) unified call center</p>

Devices/Products photo:		
		
Authorized Representative Details	AR name:	Becton Dickinson BV Saudi Limited Company
	Assigned Contact Person:	Nawaf Alsalmi
	Mobile/Phone:	+(966) 566898822
	Email:	Nawaf.M.Al.Salmie@bd.com

