

Safety Communication

رسالة سلامة

May Be Susceptible to PrintNightmare Vulnerability

Device/ Product Description:	Various Clinical laboratory Analyzers	
Manufacturer:	BD Biosciences	
Problem:	<p>The manufacturer states that it is aware of and is currently monitoring a remote code execution vulnerability in the Microsoft Windows Print spooler service, known as PrintNightmare, affecting all versions of Microsoft Windows operating systems. CVE-2021-34527 is a remote code execution vulnerability that exists when the Windows Print Spooler service improperly performs privileged file operations. An attacker who successfully exploits this vulnerability could run arbitrary code with SYSTEM privileges. An attacker could then install programs; view, change, or delete data; or create new accounts with full user rights.</p> <p>For more information, please check Here.</p> <p>If you think you had a problem with your device or a device your patient uses, please report the problem to SFDA through: NCMDR Vigilance system 19999 unified call center</p>	
Authorized Representative Details	AR name:	Becton Dickinson B.V.
	Assigned Contact Person:	Nawaf Alsalmi
	Mobile/Phone:	+(966) 566898822
	Email:	GMB-SAU-REGULATORY@bd.com